
Midmark M11 - v1.0.5
Total Cycles: 3763

Name: RITTER2
Sterilizer ID: V2313228
07/10/2024 11:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.0 F 0.0
2:00 147.6 F 1.5
4:00 149.6 F 4.0
6:00 182.7 F 7.8
8:00 200.6 F 11.6
10:00 215.4 F 4.7
12:00 225.5 F 9.2
14:00 234.4 F 11.7
16:00 237.7 F 10.9
18:00 235.5 F 9.1
20:00 248.6 F 15.2
22:00 259.1 F 21.2
24:00 269.1 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.4
1:00 272.7 F 30.2
2:00 272.3 F 29.7
3:00 271.1 F 29.3
4:00 272.1 F 30.1

Min 271.1 F 29.2
Max 273.1 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 215.4 F 0.0
10:00 213.7 F 0.0
15:00 217.1 F 0.0
20:00 210.8 F 0.0
25:00 216.9 F 0.0
30:00 218.5 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 25:12
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:03:59
CYCLE COMPLETE

07/10/2024 12:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____