
Midmark M11 - v1.0.0
Total Cycles: 1085

Name: RITTER 1
Sterilizer ID: V2178198
08/10/2021 09:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.6 F 0.0
2:00 185.4 F 2.5
4:00 186.8 F 6.5
6:00 208.9 F 11.6
8:00 223.8 F 5.8
10:00 233.8 F 10.7
12:00 231.9 F 7.3
14:00 245.4 F 13.3
16:00 243.6 F 11.8
18:00 255.7 F 18.2
20:00 266.6 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.3
1:00 272.3 F 28.7
2:00 272.3 F 28.8
3:00 272.1 F 28.4
4:00 272.5 F 29.0

Min 271.6 F 28.2
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 232.9 F 0.0
10:00 194.2 F 0.0
15:00 191.9 F 0.0
20:00 181.0 F 0.0
25:00 198.9 F 0.0
30:00 190.6 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 21:32
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:00:14
CYCLE COMPLETE

08/10/2021 10:34 AM
ACCEPTED: X
REJECTED:
Operator: U01