
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002356
11:17 05/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:51
STERILIZING 4:01
136.4 C 224kPa 4:01
136.4 C 225kPa 4:31
136.4 C 225kPa 5:01
136.4 C 225kPa 5:31
136.4 C 223kPa 6:01
136.2 C 224kPa 6:31
136.3 C 224kPa 7:01
136.1 C 220kPa 7:31
136.1 C 220kPa 7:31
VENTING 7:31
AIR DRYING 8:22
CYCLE COMPLETE 8:36

Digital Signature #
AAEF72867BB32AC1
