

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001673  
10:54 04/01/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:49  
PRESSURIZING 2:54  
STERILIZING 3:56  
136.4 C 222kPa 3:56  
136.3 C 224kPa 4:26  
136.2 C 222kPa 4:56  
136.4 C 225kPa 5:26  
136.3 C 223kPa 5:56  
136.4 C 223kPa 6:26  
136.2 C 223kPa 6:56  
136.4 C 225kPa 7:26  
136.4 C 225kPa 7:26  
VENTING 7:26  
AIR DRYING 8:17  
CYCLE COMPLETE 8:31  
-----

Digital Signature #  
1F8CFDF44907E5AB  
-----