
Midmark M11 - v1.0.0
Total Cycles: 4321

Name: RITTER 1
Sterilizer ID: V2178198
02/22/2024 11:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.7 F 0.0
2:00 177.1 F 2.1
4:00 179.8 F 5.4
6:00 186.7 F 9.1
8:00 204.4 F 13.3
10:00 217.7 F 6.1
12:00 227.3 F 10.3
14:00 231.7 F 8.1
16:00 239.8 F 11.8
18:00 238.4 F 10.0
20:00 251.0 F 16.0
22:00 261.7 F 22.2
24:00 271.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.4 F 29.4
2:00 272.0 F 28.9
3:00 271.1 F 28.5
4:00 272.1 F 29.4

Min 271.1 F 28.4
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 212.2 F 0.0
10:00 215.8 F 0.0
15:00 211.5 F 0.0
20:00 207.0 F 0.0
25:00 205.7 F 0.0
30:00 205.2 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 24:39
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:03:35
CYCLE COMPLETE

02/22/2024 12:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____