
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
2.0uS / 1.2ppm
CYCLE NUMBER 003732
9:32 31/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 2:44
STERILIZING 3:52
136.0 C 220kPa 3:52
136.2 C 222kPa 4:22
136.4 C 222kPa 4:52
136.3 C 222kPa 5:22
136.2 C 222kPa 5:52
136.4 C 223kPa 6:22
136.1 C 219kPa 6:52
136.4 C 223kPa 7:22
136.4 C 223kPa 7:22
VENTING 7:22
AIR DRYING 8:56
CYCLE COMPLETE 9:21

Digital Signature #
68C7C624846C53F6
