
Midmark M11 - v1.0.5
Total Cycles: 2400

Name: RITTER2
Sterilizer ID: V2313228
05/23/2023 12:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.2 F 0.0
2:00 182.0 F 3.6
4:00 185.7 F 8.2
6:00 207.0 F 13.4
8:00 223.2 F 8.5
10:00 233.0 F 10.9
12:00 241.5 F 11.8
14:00 240.1 F 10.6
16:00 253.7 F 17.6
18:00 265.2 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.7 F 30.1
2:00 273.5 F 30.2
3:00 272.3 F 29.8
4:00 272.3 F 29.5

Min 271.6 F 29.1
Max 273.9 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.2 F 0.0
5:00 225.5 F 0.0
10:00 211.9 F 0.0
15:00 208.7 F 0.0
20:00 208.5 F 0.0
25:00 217.8 F 0.0
30:00 233.5 F 0.0

Drying Complete

FILLING: 1:58
HEATING: 19:55
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:05
CYCLE COMPLETE

05/23/2023 01:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____