
Midmark M11 - v1.0.0
Total Cycles: 790

Name: RITTER 1
Sterilizer ID: V2178198
04/21/2021 08:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 210.1 F 0.0
2:00 189.3 F 1.9
4:00 193.1 F 6.3
6:00 207.2 F 12.1
8:00 226.9 F 7.1
10:00 238.3 F 12.7
12:00 240.3 F 10.4
14:00 236.0 F 8.6
16:00 251.3 F 15.7
18:00 263.1 F 22.4
20:00 271.8 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.6 F 28.6
2:00 272.3 F 28.4
3:00 272.1 F 28.4
4:00 272.1 F 28.4

Min 270.9 F 27.9
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.7 F 0.0
5:00 234.1 F 0.0
10:00 219.7 F 0.0
15:00 223.3 F 0.0
20:00 166.8 F 0.0
25:00 192.9 F 0.0
Dry Cycle Aborted

FILLING: 1:35
HEATING: 20:02
STERILIZING: 4:00
VENTING: 3:04
DRYING: 26:36
TOTAL CYCLE: 00:55:16

04/21/2021 09:07 AM
ACCEPTED: X
REJECTED:
Operator: U01