
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001066
15:15 06/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:34
PRESSURIZING 1:39
STERILIZING 2:42
136.0 C 222kPa 2:42
136.4 C 226kPa 3:12
136.4 C 225kPa 3:42
136.4 C 226kPa 4:12
136.2 C 224kPa 4:42
136.4 C 224kPa 5:12
136.4 C 226kPa 5:42
136.4 C 224kPa 6:12
136.4 C 224kPa 6:12
VENTING 6:12
AIR DRYING 7:03
CYCLE COMPLETE 22:03

Digital Signature #
1213CBC6F7B838B7
