

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 003926  
11:03 04/07/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:42  
PRESSURIZING 2:49  
STERILIZING 3:55  
136.0 C 220kPa 3:55  
136.4 C 223kPa 4:25  
135.8 C 217kPa 4:55  
136.4 C 221kPa 5:25  
136.3 C 222kPa 5:55  
136.4 C 224kPa 6:25  
136.1 C 222kPa 6:55  
136.4 C 224kPa 7:25  
136.4 C 224kPa 7:25  
VENTING 7:25  
AIR DRYING 8:13  
CYCLE COMPLETE 8:31

-----  
Digital Signature #  
6E3470B1405CC918  
-----