
Midmark M11 - v1.0.5
Total Cycles: 125

Name: RITTER2
Sterilizer ID: V2313228
06/22/2021 09:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 100.9 F 0.0
2:00 101.6 F 0.6
4:00 121.1 F 2.1
6:00 148.5 F 4.7
8:00 174.6 F 8.2
10:00 193.2 F 12.0
12:00 205.9 F 4.7
14:00 218.4 F 8.5
16:00 229.3 F 12.6
18:00 231.3 F 9.1
20:00 238.9 F 11.4
22:00 242.8 F 12.7
24:00 253.4 F 18.4
26:00 263.7 F 24.5
28:00 272.4 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.7
1:00 272.6 F 30.3
2:00 271.4 F 29.6
3:00 272.4 F 30.4
4:00 271.8 F 29.8

Min 271.3 F 29.6
Max 272.7 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.9 F 0.0
5:00 207.3 F 0.0
10:00 207.9 F 0.0
15:00 206.3 F 0.0
20:00 198.4 F 0.0
25:00 194.4 F 0.0
30:00 190.8 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 28:14
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:06:47
CYCLE COMPLETE

06/22/2021 10:52 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____