
Midmark M11 - v1.0.5
Total Cycles: 770

Name: RITTER2
Sterilizer ID: V2313228
02/08/2022 12:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.8 F 0.0
2:00 169.8 F 3.6
4:00 186.3 F 8.1
6:00 207.5 F 13.3
8:00 224.1 F 8.9
10:00 234.4 F 9.9
12:00 241.7 F 12.8
14:00 242.0 F 12.0
16:00 255.3 F 18.9
18:00 267.2 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.4 F 29.6
2:00 271.6 F 29.5
3:00 272.5 F 29.6
4:00 271.5 F 29.1

Min 271.4 F 29.1
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.2 F 0.0
5:00 210.2 F 0.0
10:00 223.9 F 0.0
15:00 221.0 F 0.0
20:00 220.9 F 0.0
25:00 223.8 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 19:23
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:08
CYCLE COMPLETE

02/08/2022 01:11 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____