

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3104

Name: RITTER 1  
Sterilizer ID: V2178198  
04/13/2023 08:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.1 F 0.0  
2:00 181.8 F 2.0  
4:00 188.9 F 5.9  
6:00 215.0 F 10.9  
8:00 223.2 F 5.9  
10:00 236.4 F 11.1  
12:00 233.5 F 7.9  
14:00 242.8 F 11.8  
16:00 245.1 F 12.9  
18:00 257.7 F 19.6  
20:00 268.8 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.5 F 29.1  
2:00 271.3 F 28.3  
3:00 272.6 F 29.2  
4:00 271.9 F 28.6  
  
Min 271.2 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.4 F 0.0  
5:00 214.7 F 0.0  
10:00 224.4 F 0.0  
15:00 227.3 F 0.0  
20:00 225.0 F 0.0  
25:00 224.3 F 0.0  
30:00 223.3 F 0.0  
Drying Complete

FILLING: 1:53  
HEATING: 21:10  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:00:18  
CYCLE COMPLETE

04/13/2023 09:27 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_