
Midmark M11 - v1.0.0
Total Cycles: 2906

Name: RITTER 1
Sterilizer ID: V2178198
02/15/2023 10:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.0 F 0.0
2:00 203.1 F 4.1
4:00 202.7 F 9.2
6:00 220.7 F 7.4
8:00 229.5 F 9.7
10:00 234.6 F 8.8
12:00 243.1 F 12.9
14:00 242.1 F 11.6
16:00 255.3 F 18.4
18:00 266.9 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.6 F 29.3
2:00 271.8 F 28.5
3:00 272.4 F 29.4
4:00 272.4 F 29.0

Min 271.5 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 223.4 F 0.0
10:00 227.8 F 0.0
15:00 237.5 F 0.0
20:00 228.1 F 0.0
25:00 223.5 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 19:31
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:37
CYCLE COMPLETE

02/15/2023 11:56 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____