

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 006752  
10:00 18/12/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:49  
PRESSURIZING 3:49  
STERILIZING 5:00  
136.4 C 221kPa 5:00  
136.2 C 223kPa 5:30  
136.4 C 222kPa 6:00  
136.4 C 223kPa 6:30  
136.4 C 222kPa 7:00  
136.2 C 223kPa 7:30  
136.4 C 223kPa 8:00  
136.4 C 223kPa 8:30  
136.4 C 223kPa 8:30  
VENTING 8:30  
AIR DRYING 9:28  
CYCLE COMPLETE 9:31  
-----

Digital Signature #  
B3A0BF4CF4E59D490  
-----