

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3311

Name: RITTER 1  
Sterilizer ID: V2178198  
05/31/2023 09:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.6 F 0.0  
2:00 187.4 F 0.1  
4:00 210.1 F 5.6  
6:00 228.2 F 11.0  
8:00 229.9 F 7.1  
10:00 242.0 F 13.1  
12:00 241.7 F 11.3  
14:00 239.7 F 10.1  
16:00 253.9 F 17.1  
18:00 265.7 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.0  
1:00 272.6 F 28.7  
2:00 271.5 F 27.9  
3:00 272.2 F 28.7  
4:00 272.3 F 28.5  
  
Min 271.3 F 27.8  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.9 F 0.0  
5:00 219.3 F 0.0  
10:00 222.6 F 0.0  
15:00 225.9 F 0.0  
20:00 222.3 F 0.0  
25:00 222.5 F 0.0  
30:00 222.0 F 0.0  
Drying Complete

FILLING: 1:35  
HEATING: 19:42  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:32  
CYCLE COMPLETE

05/31/2023 10:36 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_