
Midmark M11 - v1.0.0
Total Cycles: 1524

Name: RITTER 1
Sterilizer ID: V2178198
01/13/2022 12:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.6 F 0.0
2:00 172.7 F 3.8
4:00 195.1 F 8.2
6:00 212.3 F 13.4
8:00 225.8 F 7.4
10:00 238.2 F 12.5
12:00 237.8 F 9.7
14:00 237.3 F 8.8
16:00 249.3 F 14.8
18:00 261.5 F 21.7
20:00 272.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.8 F 29.0
2:00 271.8 F 28.3
3:00 271.8 F 28.7
4:00 272.7 F 29.0

Min 271.2 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.8 F 0.0
5:00 236.6 F 0.0
10:00 205.2 F 0.0
15:00 203.7 F 0.0
20:00 207.3 F 0.0
25:00 223.9 F 0.0
30:00 180.3 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 20:28
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:30
CYCLE COMPLETE

01/13/2022 01:58 PM
ACCEPTED: X
REJECTED:
Operator: U02