

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4088

Name: RITTER2  
Sterilizer ID: V2313228  
11/05/2024 03:38 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 143.7 F 0.0  
2:00 144.4 F 1.9  
4:00 165.8 F 5.3  
6:00 187.3 F 9.5  
8:00 218.1 F 8.9  
10:00 226.5 F 7.8  
12:00 237.1 F 12.5  
14:00 236.1 F 9.5  
16:00 237.9 F 9.6  
18:00 247.3 F 14.3  
20:00 259.0 F 20.7  
22:00 269.7 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.0  
1:00 273.0 F 30.0  
2:00 272.4 F 30.0  
3:00 272.2 F 29.9  
4:00 271.9 F 29.6

Min 271.0 F 28.8  
Max 273.3 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.7 F 0.0  
5:00 194.4 F 0.0  
10:00 191.9 F 0.0  
15:00 189.8 F 0.0  
20:00 187.2 F 0.0  
25:00 184.5 F 0.0  
30:00 182.6 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 23:02  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:01:49  
CYCLE COMPLETE

11/05/2024 04:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_