
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001524
10:42 23/10/2023
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 4:49
STERILIZING 5:39
136.4 C 226kPa 5:39
136.2 C 226kPa 6:09
136.4 C 227kPa 6:39
136.4 C 227kPa 7:09
136.2 C 223kPa 7:39
136.3 C 226kPa 8:09
136.4 C 226kPa 8:39
136.4 C 223kPa 9:09
136.3 C 227kPa 9:39
136.3 C 226kPa 10:09
136.2 C 225kPa 10:39
136.3 C 225kPa 11:09
136.2 C 225kPa 11:39
136.4 C 227kPa 12:09
136.4 C 226kPa 12:39
136.2 C 224kPa 13:09
136.4 C 226kPa 13:39
136.4 C 227kPa 14:09
136.2 C 225kPa 14:39
136.3 C 225kPa 15:09
136.4 C 227kPa 15:39
136.4 C 227kPa 15:39
VENTING 15:39
AIR DRYING 16:35
CYCLE COMPLETE 16:41

Digital Signature #
508CD0E70426705A
