
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000054
7:50 29/01/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:08
PRESSURIZING 4:15
STERILIZING 5:41
136.4 C 221kPa 5:41
136.4 C 223kPa 6:11
136.4 C 222kPa 6:41
136.4 C 224kPa 7:11
136.4 C 224kPa 7:41
136.0 C 222kPa 8:11
136.3 C 221kPa 8:41
136.3 C 222kPa 9:11
136.3 C 222kPa 9:11
VENTING 9:11
AIR DRYING 10:11
CYCLE COMPLETE 10:13

Digital Signature #
5EC7F67462B5F3B5
