
Midmark M11 - v1.0.0
Total Cycles: 2061

Name: RITTER 1
Sterilizer ID: V2178198
06/27/2022 12:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.7 F 0.0
2:00 192.0 F 4.4
4:00 213.0 F 9.4
6:00 229.3 F 6.8
8:00 237.1 F 10.6
10:00 233.7 F 7.8
12:00 242.1 F 10.9
14:00 246.6 F 13.6
16:00 259.7 F 20.6
18:00 271.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 28.9
2:00 271.3 F 28.2
3:00 272.6 F 29.1
4:00 272.0 F 28.5

Min 271.3 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 239.7 F 0.0
10:00 232.8 F 0.0
15:00 226.2 F 0.0
20:00 218.8 F 0.0
25:00 219.0 F 0.0
30:00 217.8 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 18:39
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:27
CYCLE COMPLETE

06/27/2022 01:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____