
Midmark M11 - v1.0.5
Total Cycles: 4085

Name: RITTER2
Sterilizer ID: V2313228
11/05/2024 11:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.8 F 0.0
2:00 154.1 F 1.0
4:00 199.2 F 3.9
6:00 184.7 F 7.6
8:00 205.8 F 11.5
10:00 211.1 F 4.6
12:00 229.8 F 8.6
14:00 239.9 F 12.8
16:00 234.8 F 9.5
18:00 239.6 F 10.5
20:00 245.2 F 13.2
22:00 255.8 F 18.9
24:00 265.5 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.6 F 29.7
2:00 272.0 F 29.1
3:00 271.6 F 29.3
4:00 272.5 F 29.7

Min 271.3 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.0 F 0.0
5:00 222.7 F 0.0
10:00 223.2 F 0.0
15:00 222.2 F 0.0
20:00 211.5 F 0.0
25:00 204.5 F 0.0
30:00 203.4 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 25:59
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:04:57
CYCLE COMPLETE

11/05/2024 12:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____