
Midmark M11 - v1.0.5
Total Cycles: 4396

Name: RITTER2
Sterilizer ID: V2313228
03/04/2025 09:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.3 F 0.0
2:00 74.2 F 0.0
4:00 122.8 F 0.0
6:00 158.3 F 0.8
8:00 142.6 F 3.7
10:00 170.5 F 6.9
12:00 185.9 F 10.6
14:00 200.5 F 7.4
16:00 211.0 F 5.7
18:00 222.3 F 9.3
20:00 232.9 F 13.4
22:00 234.5 F 9.7
24:00 238.9 F 10.1
26:00 244.9 F 13.3
28:00 256.0 F 19.1
30:00 266.1 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.2
1:00 272.5 F 29.4
2:00 272.8 F 29.9
3:00 271.9 F 29.3
4:00 272.4 F 29.4

Min 271.7 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.0 F 0.0
5:00 199.9 F 0.0
10:00 205.5 F 0.0
15:00 207.3 F 0.0
20:00 207.7 F 0.0
25:00 211.4 F 0.0
30:00 209.6 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 31:42
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:10:11
CYCLE COMPLETE

03/04/2025 10:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____