
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001518
7:25 10/21/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:04
PRESSURIZING 6:58
STERILIZING 8:35
136.0 C 220kPa 8:35
136.1 C 218kPa 9:06
136.3 C 221kPa 9:36
136.3 C 219kPa 10:06
136.1 C 219kPa 10:36
136.1 C 219kPa 11:06
136.4 C 221kPa 11:36
136.5 C 222kPa 12:05
VENTING 12:06
AIR DRYING 13:16
CYCLE COMPLETE 43:16

Digital Signature #
C7EA5E5DC6BDE288
