
Midmark M11 - v1.0.0
Total Cycles: 3812

Name: RITTER 1
Sterilizer ID: V2178198
09/25/2023 03:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.0 F 0.0
2:00 176.7 F 2.6
4:00 195.3 F 6.9
6:00 213.2 F 11.5
8:00 228.4 F 6.6
10:00 234.8 F 11.6
12:00 235.7 F 8.6
14:00 242.2 F 11.0
16:00 245.9 F 12.9
18:00 257.4 F 19.0
20:00 267.5 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 28.5
1:00 271.7 F 28.5
2:00 271.9 F 28.8
3:00 272.2 F 28.9
4:00 272.3 F 28.9

Min 271.4 F 28.4
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.8 F 0.0
5:00 228.9 F 0.0
10:00 221.1 F 0.0
15:00 209.5 F 0.0
20:00 200.2 F 0.0
25:00 206.4 F 0.0
30:00 201.3 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 21:52
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:01:19
CYCLE COMPLETE

09/25/2023 04:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____