
Midmark M11 - v1.0.5
Total Cycles: 36

Name: RITTER2
Sterilizer ID: V2313228
05/20/2021 09:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.6 F 0.0
2:00 158.2 F 2.2
4:00 164.8 F 5.8
6:00 186.9 F 10.3
8:00 210.9 F 5.3
10:00 222.6 F 9.2
12:00 233.2 F 10.2
14:00 238.7 F 12.1
16:00 240.4 F 11.2
18:00 252.8 F 17.6
20:00 263.9 F 24.2
22:00 271.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.7 F 30.1
2:00 271.7 F 29.3
3:00 271.8 F 30.1
4:00 272.4 F 29.9

Min 271.3 F 29.2
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.5 F 0.0
5:00 208.7 F 0.0
10:00 208.2 F 0.0
15:00 201.3 F 0.0
20:00 196.8 F 0.0
25:00 196.3 F 0.0
30:00 195.2 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 22:08
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:46
CYCLE COMPLETE

05/20/2021 10:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____