
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007016
7:25 01/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:31
PRESSURIZING 3:32
STERILIZING 4:21
136.0 C 220kPa 4:21
136.4 C 224kPa 4:51
136.3 C 221kPa 5:21
136.4 C 220kPa 5:51
136.2 C 221kPa 6:21
136.2 C 220kPa 6:51
135.9 C 219kPa 7:21
136.3 C 222kPa 7:51
136.3 C 222kPa 7:51
VENTING 7:51
AIR DRYING 8:39
CYCLE COMPLETE 8:43

Digital Signature #
FF4B4C8521170E1E
