
Midmark M11 - v1.0.0
Total Cycles: 1947

Name: RITTER 1
Sterilizer ID: V2178198
05/18/2022 02:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.2 F 0.0
2:00 184.0 F 4.5
4:00 201.0 F 9.2
6:00 217.7 F 7.8
8:00 231.4 F 9.4
10:00 235.7 F 8.9
12:00 242.0 F 12.4
14:00 241.1 F 10.9
16:00 254.8 F 17.8
18:00 266.4 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 28.9
2:00 271.9 F 28.8
3:00 272.4 F 28.8
4:00 271.8 F 28.8

Min 271.5 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.9 F 0.0
5:00 230.1 F 0.0
10:00 229.2 F 0.0
15:00 228.7 F 0.0
20:00 232.2 F 0.0
25:00 225.4 F 0.0
30:00 222.5 F 0.0

Drying Complete

FILLING: 1:43
HEATING: 19:33
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:27
CYCLE COMPLETE

05/18/2022 03:46 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____