
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000736
9:41 30/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:22
PRESSURIZING 1:26
STERILIZING 2:27
136.0 C 220kPa 2:27
136.4 C 223kPa 2:57
136.4 C 224kPa 3:27
136.4 C 221kPa 3:57
136.4 C 225kPa 4:27
136.3 C 222kPa 4:57
136.4 C 224kPa 5:27
136.3 C 224kPa 5:57
136.3 C 224kPa 5:57
VENTING 5:57
AIR DRYING 6:47
CYCLE COMPLETE 6:58

Digital Signature #
08FCE4A78B166180
