
Midmark M11 - v1.0.5
Total Cycles: 3043

Name: RITTER2
Sterilizer ID: V2313228
10/23/2023 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.7 F 0.0
2:00 138.2 F 1.8
4:00 157.0 F 5.0
6:00 180.9 F 8.9
8:00 202.6 F 13.4
10:00 214.7 F 6.7
12:00 227.7 F 11.3
14:00 230.4 F 7.9
16:00 242.6 F 12.8
18:00 243.2 F 12.5
20:00 255.4 F 18.9
22:00 266.5 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.2
1:00 271.8 F 29.7
2:00 272.5 F 29.8
3:00 271.5 F 29.2
4:00 271.5 F 29.7

Min 271.2 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.2 F 0.0
5:00 206.9 F 0.0
10:00 226.9 F 0.0
15:00 223.5 F 0.0
20:00 226.4 F 0.0
25:00 232.7 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 23:29
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:02:03
CYCLE COMPLETE

10/23/2023 01:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____