
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000182
9:26 17/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:26
PRESSURIZING 4:26
STERILIZING 5:36
136.0 C 221kPa 5:36
136.4 C 226kPa 6:06
136.4 C 225kPa 6:36
136.3 C 225kPa 7:06
136.4 C 226kPa 7:36
136.4 C 226kPa 8:06
136.4 C 224kPa 8:36
136.4 C 226kPa 9:06
136.4 C 226kPa 9:06
VENTING 9:06
AIR DRYING 10:07
CYCLE COMPLETE 25:07

Digital Signature #
BB2E857F8A1B23BD
