
Midmark M11 - v1.0.0
Total Cycles: 1643

Name: RITTER 1
Sterilizer ID: V2178198
02/16/2022 02:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.3 F 0.0
2:00 183.9 F 3.3
4:00 190.6 F 7.6
6:00 212.7 F 13.4
8:00 225.7 F 7.9
10:00 238.9 F 13.2
12:00 239.1 F 10.7
14:00 237.3 F 9.0
16:00 251.6 F 16.0
18:00 263.5 F 23.0
20:00 271.5 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.6 F 29.0
2:00 271.5 F 28.2
3:00 272.6 F 29.1
4:00 272.2 F 28.6

Min 271.4 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 226.9 F 0.0
10:00 249.3 F 0.0
15:00 236.1 F 0.0
20:00 223.4 F 0.0
25:00 219.8 F 0.0
30:00 220.9 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 20:04
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

02/16/2022 03:18 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____