
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004426
13:45 22/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:31
STERILIZING 3:25
136.5 C 220kPa 3:25
136.4 C 225kPa 3:55
136.3 C 223kPa 4:25
136.4 C 223kPa 4:55
136.4 C 224kPa 5:25
136.1 C 221kPa 5:55
136.4 C 220kPa 6:25
136.3 C 220kPa 6:55
136.3 C 220kPa 6:55
VENTING 6:55
AIR DRYING 7:47
CYCLE COMPLETE 7:50

Digital Signature #
9641AFD33E20C885
