
Midmark M11 - v1.0.5
Total Cycles: 2100

Name: RITTER2
Sterilizer ID: V2313228
02/28/2023 12:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.7 F 0.0
2:00 133.2 F 0.1
4:00 150.6 F 1.2
6:00 181.7 F 5.4
8:00 207.6 F 9.9
10:00 220.1 F 6.8
12:00 228.6 F 10.6
14:00 230.6 F 7.6
16:00 241.9 F 12.2
18:00 243.8 F 12.9
20:00 256.4 F 19.5
22:00 267.6 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.7 F 30.0
2:00 272.0 F 29.3
3:00 271.7 F 29.7
4:00 272.7 F 30.0

Min 271.3 F 29.1
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.4 F 0.0
5:00 218.3 F 0.0
10:00 233.7 F 0.0
15:00 215.1 F 0.0
20:00 213.1 F 0.0
25:00 219.5 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 23:25
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:01:34
CYCLE COMPLETE

02/28/2023 01:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____