
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000865
10:56 14/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:26
PRESSURIZING 3:31
STERILIZING 4:53
136.0 C 226kPa 4:53
135.9 C 229kPa 5:23
136.4 C 232kPa 5:53
136.2 C 232kPa 6:23
136.2 C 231kPa 6:53
136.3 C 231kPa 7:23
136.3 C 230kPa 7:53
136.3 C 230kPa 8:23
136.3 C 230kPa 8:23
VENTING 8:23
AIR DRYING 9:17
CYCLE COMPLETE 24:17

Digital Signature #
92265816D2DFEBC5
