
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001959
8:17 16/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:10
PRESSURIZING 2:16
STERILIZING 3:23
136.4 C 224kPa 3:23
136.3 C 224kPa 3:53
136.4 C 221kPa 4:23
136.4 C 225kPa 4:53
136.1 C 223kPa 5:23
136.4 C 225kPa 5:53
136.3 C 223kPa 6:23
136.4 C 225kPa 6:53
136.4 C 225kPa 6:53
VENTING 6:53
AIR DRYING 7:45
CYCLE COMPLETE 19:05

Digital Signature #
A067B72646C4D3C2
