

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4180

Name: RITTER2  
Sterilizer ID: V2313228  
12/04/2024 10:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.9 F 0.0  
2:00 73.4 F 0.0  
4:00 107.7 F 0.0  
6:00 104.8 F 0.1  
8:00 149.2 F 0.3  
10:00 174.6 F 0.5  
12:00 201.8 F 1.3  
14:00 213.2 F 4.5  
16:00 222.9 F 7.7  
18:00 232.0 F 11.3  
20:00 231.0 F 7.1  
22:00 241.5 F 12.5  
24:00 239.9 F 10.7  
26:00 237.6 F 9.6  
28:00 251.4 F 16.2  
30:00 262.8 F 22.8  
32:00 272.8 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.7 F 29.3  
2:00 271.6 F 28.5  
3:00 272.7 F 29.4  
4:00 272.0 F 28.7  
  
Min 271.5 F 28.4  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.3 F 0.0  
5:00 191.1 F 0.0  
10:00 200.7 F 0.0  
15:00 198.1 F 0.0  
20:00 200.4 F 0.0  
25:00 208.4 F 0.0  
30:00 211.3 F 0.0  
  
Drying Complete

FILLING: 0:55  
HEATING: 32:20  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:10:43  
CYCLE COMPLETE

12/04/2024 11:23 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_