
Midmark M11 - v1.0.5
Total Cycles: 3360

Name: RITTER2
Sterilizer ID: V2313228
02/06/2024 02:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.6 F 0.0
2:00 100.2 F 0.0
4:00 120.4 F 0.7
6:00 148.1 F 3.6
8:00 177.2 F 7.4
10:00 210.1 F 11.4
12:00 215.5 F 5.2
14:00 227.2 F 9.7
16:00 233.5 F 9.5
18:00 239.7 F 11.6
20:00 238.0 F 9.9
22:00 251.5 F 16.5
24:00 263.1 F 23.3
26:00 272.0 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.5 F 29.5
2:00 271.3 F 28.8
3:00 272.8 F 29.6
4:00 271.7 F 28.9

Min 271.2 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.6 F 0.0
5:00 217.2 F 0.0
10:00 199.9 F 0.0
15:00 198.2 F 0.0
20:00 195.4 F 0.0
25:00 193.5 F 0.0
30:00 191.8 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 26:10
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:04:18
CYCLE COMPLETE

02/06/2024 03:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____