

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002062  
7:16 19/07/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:54  
PRESSURIZING 1:59  
STERILIZING 2:59  
136.0 C 220kPa 2:59  
136.2 C 221kPa 3:29  
136.4 C 220kPa 3:59  
136.4 C 220kPa 4:29  
136.4 C 223kPa 4:59  
136.1 C 218kPa 5:29  
135.9 C 219kPa 5:59  
136.3 C 222kPa 6:29  
136.3 C 222kPa 6:29  
VENTING 6:30  
AIR DRYING 7:21  
CYCLE COMPLETE 7:24

-----  
Digital Signature #  
5ADEB00D8F3066F8  
-----