
Midmark M11 - v1.0.0
Total Cycles: 1326

Name: RITTER 1
Sterilizer ID: V2178198
11/01/2021 09:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.4 F 0.0
2:00 187.1 F 2.0
4:00 206.0 F 6.4
6:00 219.9 F 11.5
8:00 225.0 F 6.3
10:00 235.7 F 11.4
12:00 235.3 F 8.4
14:00 242.5 F 11.0
16:00 245.4 F 13.2
18:00 257.4 F 19.6
20:00 268.2 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.7 F 29.0
2:00 272.1 F 28.9
3:00 272.5 F 28.9
4:00 271.9 F 28.8

Min 271.7 F 28.5
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.3 F 0.0
5:00 217.3 F 0.0
10:00 225.4 F 0.0
15:00 217.9 F 0.0
20:00 205.5 F 0.0
25:00 202.4 F 0.0
30:00 210.5 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 21:15
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:00:13
CYCLE COMPLETE

11/01/2021 10:31 AM
ACCEPTED: X
REJECTED:
Operator: U01