
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000680
16:07 21/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:42
STERILIZING 3:42
136.0 C 220kPa 3:42
136.4 C 223kPa 4:12
136.4 C 223kPa 4:42
136.4 C 223kPa 5:12
136.4 C 224kPa 5:42
136.3 C 224kPa 6:12
136.2 C 217kPa 6:42
136.3 C 225kPa 7:12
136.3 C 225kPa 7:12
VENTING 7:12
AIR DRYING 8:14
CYCLE COMPLETE 8:17

Digital Signature #
5DC7D0857CC0C17B
