
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000645
13:47 10/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 2:30
STOP BUTTON PRESSED
NOT STERILE
END TIME 2:30

Digital Signature #
04542ADF5DB6061D
