
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005969
11:39 13/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:02
STERILIZING 3:03
136.4 C 220kPa 3:03
136.4 C 225kPa 3:33
136.4 C 222kPa 4:03
136.4 C 224kPa 4:33
136.4 C 222kPa 5:03
136.4 C 222kPa 5:33
136.3 C 220kPa 6:03
136.4 C 223kPa 6:33
136.4 C 223kPa 6:33
VENTING 6:33
AIR DRYING 7:23
CYCLE COMPLETE 37:23

Digital Signature #
D777C9ABDA4BCAB4
