
Midmark M11 - v1.0.0
Total Cycles: 4698

Name: RITTER 1
Sterilizer ID: V2178198
07/05/2024 08:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.0 F 0.0
2:00 72.8 F 0.3
4:00 81.6 F 1.3
6:00 109.7 F 3.3
8:00 139.0 F 5.8
10:00 162.8 F 8.9
12:00 181.3 F 12.4
14:00 201.5 F 3.4
16:00 214.5 F 6.8
18:00 224.6 F 10.3
20:00 229.8 F 8.7
22:00 235.6 F 9.9
24:00 237.4 F 9.2
26:00 244.6 F 12.8
28:00 256.3 F 18.9
30:00 266.7 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.5
1:00 272.4 F 28.9
2:00 272.6 F 29.2
3:00 272.3 F 29.1
4:00 271.7 F 28.6

Min 271.5 F 28.4
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.1 F 0.0
5:00 199.8 F 0.0
10:00 216.5 F 0.0
15:00 224.5 F 0.0
20:00 230.0 F 0.0
25:00 229.3 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 31:25
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:09:38
CYCLE COMPLETE

07/05/2024 09:20 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____