
Midmark M11 - v1.0.0
Total Cycles: 4569

Name: RITTER 1
Sterilizer ID: V2178198
05/21/2024 02:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 86.3 F 0.0
2:00 87.9 F 0.6
4:00 108.2 F 2.3
6:00 140.3 F 4.5
8:00 165.5 F 7.2
10:00 182.3 F 10.3
12:00 210.3 F 8.2
14:00 223.9 F 6.2
16:00 228.0 F 9.7
18:00 241.0 F 12.3
20:00 235.7 F 9.2
22:00 240.9 F 11.3
24:00 243.0 F 11.8
26:00 253.9 F 17.4
28:00 263.9 F 23.3
30:00 272.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.1 F 29.1
2:00 272.8 F 29.4
3:00 271.1 F 28.9
4:00 272.2 F 29.2

Min 271.1 F 28.8
Max 273.1 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.6 F 0.0
5:00 231.7 F 0.0
10:00 226.9 F 0.0
15:00 221.1 F 0.0
20:00 237.0 F 0.0
25:00 226.9 F 0.0
30:00 224.7 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 31:00
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:09:34
CYCLE COMPLETE

05/21/2024 03:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____