
Midmark M11 - v1.0.0
Total Cycles: 5350

Name: RITTER 1
Sterilizer ID: V2178198
02/20/2025 03:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.6 F 0.0
2:00 125.6 F 1.3
4:00 140.6 F 3.0
6:00 173.0 F 5.0
8:00 194.0 F 6.7
10:00 209.2 F 8.5
12:00 229.8 F 11.4
14:00 230.9 F 7.3
16:00 236.2 F 9.8
18:00 245.5 F 13.5
20:00 242.2 F 11.3
22:00 238.0 F 9.3
24:00 250.3 F 15.1
26:00 259.6 F 20.3
28:00 268.0 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 271.5 F 28.1
2:00 271.8 F 28.4
3:00 272.2 F 28.7
4:00 272.6 F 28.9

Min 271.2 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.7 F 0.0
5:00 232.0 F 0.0
10:00 222.6 F 0.0
15:00 224.2 F 0.0
20:00 215.8 F 0.0
25:00 212.9 F 0.0
30:00 211.1 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:44
STERILIZING: 4:00
VENTING: 3:53
DRYING: 30:00
TOTAL CYCLE: 01:08:31
CYCLE COMPLETE

02/20/2025 04:29 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____