
Midmark M11 - v1.0.0
Total Cycles: 2081

Name: RITTER 1
Sterilizer ID: V2178198
07/04/2022 10:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.2 F 0.0
2:00 162.3 F 2.1
4:00 188.7 F 5.4
6:00 196.9 F 10.2
8:00 216.1 F 4.9
10:00 228.1 F 8.5
12:00 240.1 F 13.4
14:00 238.4 F 10.4
16:00 237.3 F 7.9
18:00 249.3 F 14.9
20:00 261.1 F 21.5
22:00 271.8 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.4 F 28.9
2:00 271.4 F 28.1
3:00 272.4 F 29.0
4:00 272.0 F 28.5

Min 271.3 F 28.1
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.9 F 0.0
5:00 215.4 F 0.0
10:00 228.3 F 0.0
15:00 234.3 F 0.0
20:00 224.8 F 0.0
25:00 227.8 F 0.0
30:00 224.9 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 22:34
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:37
CYCLE COMPLETE

07/04/2022 11:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____