

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000390  
9:35 25/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CYCLE ABORTED 4:16  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 4:16  
-----

Digital Signature #  
3A1CFA7E5FBD0CED  
-----