
Midmark M11 - v1.0.0
Total Cycles: 2213

Name: RITTER 1
Sterilizer ID: V2178198
08/09/2022 08:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.6 F 0.0
2:00 172.8 F 1.0
4:00 175.0 F 5.0
6:00 210.5 F 9.5
8:00 224.4 F 6.2
10:00 223.7 F 10.7
12:00 230.2 F 6.9
14:00 240.5 F 13.1
16:00 240.7 F 11.0
18:00 252.4 F 17.4
20:00 264.7 F 24.1
22:00 271.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 271.6 F 29.2
2:00 272.2 F 29.7
3:00 272.7 F 30.0
4:00 272.8 F 29.9

Min 271.2 F 28.7
Max 273.3 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 210.1 F 0.0
10:00 210.6 F 0.0
15:00 211.2 F 0.0
20:00 233.9 F 0.0
25:00 230.7 F 0.0
30:00 230.4 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 22:09
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 01:00:01
CYCLE COMPLETE

08/09/2022 09:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____