

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001833  
8:50 22/06/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:23  
PRESSURIZING 3:27  
STERILIZING 4:35  
136.1 C 220kPa 4:35  
136.4 C 224kPa 5:05  
136.2 C 219kPa 5:35  
136.4 C 222kPa 6:05  
136.1 C 223kPa 6:35  
136.4 C 223kPa 7:05  
136.2 C 222kPa 7:35  
136.4 C 222kPa 8:05  
136.4 C 222kPa 8:05  
VENTING 8:05  
AIR DRYING 9:01  
CYCLE COMPLETE 9:13

-----  
Digital Signature #  
BA9402679EDB548D  
-----